

Ricardo Morren, General Manager
Juvo Solutions Pty Ltd
21A Tacoma Circuit,
Canning Vale
WA 6155

18th May 2020

Dear Ricardo,

Reference: testing the following Juvo Solutions Pty Ltd products:

- JH6160 Transfer bench (Also known as 1018J, **1110** & CR1597)
- JH6160 Padded transfer bench (Also known as 1018J & **1111**)

to 140 kg user weight, our Laboratory reference is 492722-2.

This letter is to advise you that this Laboratory has finished testing your products to the following Standard method:

- AS/NZS 4069:2013 Bath boards and bath seats – Product requirements

The official eight (8) page report, dated 18th May 2020 and referencing job number 492722-2 identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Juvo Solutions Pty Ltd JH6160 (Also known as 1018J, **1110** & CR1597) Transfer bench and JH6160 (Also known as 1018J & **1111**) Padded transfer bench, as represented by the test samples and detailed in the test report, comply with the relevant requirements of AS/NZS 4069:2013 Bath boards and bath seats – Product requirements, conditional that the following labelling requirements are met:

- Clause 9(b) - The test sample was labelled 130 kg but tested to 140 kg at your request. If claiming compliance to 140 kg (as tested) a revised product label displaying this user weight (140 kg) will need to be attached to all products made available for sale.
- Clause 9(c & d) - The test sample did not have a unique serial number or date of manufacture attached. For compliance to this standard a label which includes a serial number/manufacture date must be attached to all products made available for sale.

NATA Accredited Laboratory No.2953
Accredited for compliance with ISO/IEC 17025- Testing

NovitaTech (a division of Novita Disability Services)
171 Days Road, Regency Park, South Australia 5010 PO
Box 2438, Regency Park, South Australia 5942
Telephone (08) 8243 8261
testing@novita.org.au • novitatech.com.au
ABN 53 137 537 636



Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
Authorised signatory

The supplier is responsible for organising the collection of tested equipment samples. Unclaimed samples will not be retained longer than three (3) months from report date.

NATA Accredited Laboratory No.2953
Accredited for compliance with ISO/IEC 17025- Testing

NovitaTech (a division of Novita Disability Services)
171 Days Road, Regency Park, South Australia 5010 PO
Box 2438, Regency Park, South Australia 5942
Telephone (08) 8243 8261
testing@novita.org.au • novitotech.com.au
ABN 53 137 537 636

