



Mark Braden
Crescent Healthcare
1 Allenby Street,
Coburg, Vic 3058

31st October 2013

Dear Mark,

Reference: testing the Crescent Healthcare Static toilet chair (Care-Assist CR1518A, Ansa 1312, Care-Quip B4002-AS) to 200 kg user weight, our Laboratory reference is 491993.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 3973:2009, AS/NZS 4688.2:2000 and AS/NZS 4688.3-2000 (except materials and surface finishes).

The official nine (9) page report, dated 31st October 2013 and referencing job number 491993 identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Crescent Healthcare Static toilet chair (Care-Assist CR1518A, Ansa 1312, Care-Quip B4002-AS) as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of the above referenced standards (excluding the methods indicated in the report as "not tested") for user mass 200 kg.

If claiming compliance to 200 kg user weight, an updated product label must be attached to all products made available for sale detailing the revised SWL (200 kg)

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
Authorised signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

