



Mark Braden
Crescent Healthcare
1 Allenby Street,
Coburg, Vic. 3058

15th June 2011

Dear Mark,

Reference: testing the Crescent Healthcare Shower chairs (150 kg) as follows:
CR1502 Care-assist Shower chair with arms – Aluminium – 510 mm
1322 ANSA Extra Care Shower chair with arms – Aluminium – 510 mm
Care-quip B4002WA-A Shower chair with arms – Aluminium – 510 mm,
our Laboratory reference is 491517-1.

This letter is to advise you that this Laboratory has finished testing your products to the following Standard method: AS/NZS 3973:2009 *Shower/toilet chairs – static*. (except corrosion resistance and surface finishes)

The official nine (9) page report, dated 15th June 2011 and referencing job number 491517-1, identifies the sample, the test documents and the testing facility. All conformances with the Australian Standards are indicated in the report.

The Crescent Healthcare Shower chair described above, as represented by the test sample, complies with the relevant requirements of the Australian Standard AS/NZS 3973:2009 (excluding the methods indicated in the report as “not tested”) for user mass 150 kg.

If claiming compliance to 150 kg user weight (as tested) a label must be permanently attached to all products made available for sale displaying the following information:

Recommended maximum user weight (150 kg)

Please do not hesitate to contact us if you have any enquiries or comments regarding this test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
NATA signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
testing@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

