

Mark Braden Crescent healthcare 1 Allenby Street Coburg Vic 3058

8 September 2009

Dear Mark,

Reference: testing the Crescent Healthcare Care-Assist CR1550, Ansa 1751, Care-Quip B4014 ASD Over Toilet Frames, our Laboratory reference is 491224.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS3973:2009 *Shower/Toilet Chairs (mobile and static)* with the exception of materials tests.

The official seven (7) page report, dated 7 September 2009 and referencing job number 491224-1, identifies the sample, the test documents and the testing facility. All conformances with the Australian standards are indicated in the report.

The Crescent Healthcare Care-Assist CR1550, Ansa 1751, Care-Quip B4014\_ASD Over Toilet Frames as represented by the test samples, including the updated documentation, with the provided labelling permanently affixed to the product as stated in the test report, complies with the relevant requirements of the Australian Standard AS3973:2009 (excluding the methods indicated in the report as "not tested") for user mass 125kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Ben Symonds BSc/BEng(Biomed) Manager, Test Laboratory NATA signatory



