

Mark Braden Crescent Healthcare 1 Allenby Street, Coburg, Vic. 3058

15<sup>th</sup> June 2011

Dear Mark,

Reference: testing the Crescent Healthcare Over toilet frames (150 kg) as follows: CR1553 Care-assist Over toilet frame – Aluminium – 510 mm 1760AL ANSA Extra Care Over toilet frame – Aluminium – 510 mm B4014WA-A Care-quip Over toilet frame – Aluminium – 510 mm, our Laboratory reference is 491518-1.

This letter is to advise you that this Laboratory has finished testing your products to the following Standard method: AS/NZS 3973:2009 *Shower/toilet chairs – static.* (except corrosion resistance and surface finishes)

The official nine (9) page report, dated 15<sup>th</sup> June 2011 and referencing job number 491518-1, identifies the sample, the test documents and the testing facility. All conformances with the Australian Standards are indicated in the report.

The Crescent Healthcare Over toilet frame described above, as represented by the test sample, complies with the relevant requirements of the Australian Standard AS/NZS 3973:2009 (excluding the methods indicated in the report as "not tested") for user mass 150 kg.

This compliance is conditional that a label/s be permanently attached to all products made available for sale displaying the following information:

- 1) Registered name / trademark & address of manufacturer or distributor
- 2) Recommended maximum user weight (150 kg)
- 3) Notice that the chair is suitable for use in showers or bathrooms
- 4) Notice that the chair is not suitable for use in swimming pools or corrosive environments.

Please do not hesitate to contact us if you have any enquiries or comments regarding this test.

Kind regards,

Wayne Wurfel Senior Test Technician, Test Laboratory NATA signatory

