



Mark Braden

Crescent Healthcare
1 Allenby Street,
Coburg, Vic. 3058

23rd September 2011

Dear Mark,

Reference: testing the Crescent Healthcare CR2064 Furniture chair (High back, standard width) our Laboratory reference is 491553.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 4688.2-2000 and AS/NZS 4688.3-2000 (except ignitibility, materials and finishes).

The official six (6) page report, dated 23rd September 2011 and referencing job number 491553, identifies the sample, the test documents and the testing facility. All conformances with the above referenced standards are indicated in the report.

The Crescent Healthcare CR2064 Furniture chair (High back, standard width) as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 4688.2-2000 and AS/NZS 4688.3-2000 (excluding the methods indicated in the report as "not tested") for user mass 300 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
(NATA signatory)

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

