



Curtis Dartnall, Product Manager
Unicare Health
29 Alexandra Place,
Bentley, WA 6102

6th October 2016

Dear Curtis,

Reference: testing the Unicare Health Bedside commode to 150 kg user weight, our Laboratory reference is 492635.

This letter is to advise you that this Laboratory has finished testing your product to the relevant clauses of the following Standards:

- AS/NZS 3973:2009 Shower / toilet chairs (mobile and static)
- AS/NZS 4688.2-2000 Furniture chairs – Determination of strength & durability
- AS/NZS 4688.3-2000 Furniture chairs – Determination of stability

The official ten (10) page report, dated 6th October 2016 and referencing job number 492635, identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standard are indicated in the report.

The Unicare Health Bedside commode, as represented by the test sample, complies with the relevant requirements of the above referenced standards (excluding the methods indicated in the report as “not tested”) for user mass 150 kg.

Please note that if you are claiming compliance to the revised user weight of 150 kg for this product an updated maximum user weight label of 150 kg must be attached to all products made available for sale.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Authorised signatory.

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

