



Curtis Dartnall, Marketing Coordinator
Unicare Health

29 Alexandra Place,
PO Box 29,
Bentley, WA 6102

23rd September 2013

Dear Curtis,

Reference: testing the Unicare Health Aluminium underarm axilla crutches (Codes 6010, 6011 & 6012) to 200 kg SWL, our Laboratory reference is 491972.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard methods:

- ISO 11334.1:2007 Assistive products for walking – Requirements & test methods
- EN 1985:1999 Walking aids – General requirements & test methods

The official four (4) page report, dated 23rd September 2013 and referencing job number 491972, identifies the sample, the test documents and the testing facility. All conformances with the above referenced standards are indicated in the report.

The Unicare Health Aluminium underarm axilla crutches (Codes 6010, 6011 & 6012) as represented by the test samples, including the supplied documentation as stated in the test report, complies with the relevant requirements of the ISO 11334.1:2007 and EN 1985:1999 Standards (excluding the methods indicated in the report as “not tested”) for user mass 200 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
Authorised Signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

