



Curtis Dartnall, Marketing Coordinator
Unicare Health

29 Alexandra Place,
PO Box 29,
Bentley, WA 6102

24th September 2013

Dear Curtis,

Reference: testing the Unicare Health Forearm (elbow) crutches, Codes 6130 & 6131 to 175 kg SWL, our Laboratory reference is 491969.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method:

- ISO 11334.1:2007 Assistive products for walking – Requirements & test methods

The official five (5) page report, dated 24th September 2013 and referencing job number 491969, identifies the sample, the test documents and the testing facility. All conformances with the above referenced standard are indicated in the report.

The Unicare Health Forearm (elbow) crutches, Codes 6130 & 6131 as represented by the test samples, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of the ISO 11334.1:2007 standards (excluding the methods indicated in the report as “not tested”) for user mass 175 kg, conditional on the following:

If claiming compliance to 175 kg user weight, an updated user weight label (175 kg) must be attached to all products made available for sale.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
Authorised Signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

