



Curtis Dartnall, Marketing Coordinator
Unicare Health
29 Alexandra Place,
PO Box 29,
Bentley, WA 6102

18th May 2017

Dear Curtis,

Reference: testing the following Unicare Health products:

- 6540 Walking frame, Duralite with wheels and gliders – Adult
- 6541 Walking frame, Duralite with wheels and gliders – Adult tall to 200 kg SWL, our Laboratory reference is 492724.

This letter is to advise you that this Laboratory has finished testing your products to the relevant requirements of following Standard :

- ISO 11199.2:2005 Walking aids manipulated by both arms – Rollators

The official six (6) page report, dated 15th May 2017 and referencing job number 492724, identifies the sample, the test documents and the testing facility. All conformances with the above referenced standards are indicated in the report.

The Unicare Health 6540 and 6541 Duralite walking frames with wheels and gliders as represented by the test samples, including the supplied documentation as stated in the test report, comply with the relevant requirements of ISO 11199.2:2005 Walking aids manipulated by both arms – Rollators, for user mass 200 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician, Test Laboratory
Authorised Signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

