

Issued to:

Unicare Health
PO Box 29
Bentley WA 6012


NovitaTech Test Laboratory
NATA Accreditation Number 2953

Certificate of Compliance

Product Name Rollator
Product Details Rollator Cruiser

The product listed above complies with the requirements of ISO 11199-2 for user weight 175 kg.

Yours sincerely,



Wayne Wurfel
Senior Test Technician
testlab@novita.org.au

Date of Issue 30/09/2013

