



Curtis Dartnall, Marketing Coordinator  
Unicare Health

29 Alexandra Place,  
PO Box 29,  
Bentley, WA 6102

30<sup>th</sup> September 2013

Dear Curtis,

Reference: testing the Unicare Health Seat walker, Cruiser Grande (Product code 6765) to 200 kg SWL, our Laboratory reference is 491981.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard methods:

- ISO 11199.2;2005 Walking aids manipulated by both arms – Rollators

The official six (6) page report, dated 30th September 2013 and referencing job number 491981, identifies the sample, the test documents and the testing facility. All conformances with the above referenced standards are indicated in the report.

The Unicare Health Seat walker, Cruiser Grande (Product code 6765) as represented by the test sample, including the supplied documentation as stated in the test report, complies with the relevant requirements of ISO 11199.2;2005 Walking aids manipulated by both arms – Rollators (excluding the methods indicated in the report as “not tested”) for user mass 200 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel  
Senior Test Technician, Test Laboratory  
Authorised Signatory

**NovitaTech (a division of Novita Children's Services)**  
171 Days Road, Regency Park, South Australia 5010  
PO Box 2438, Regency Park, South Australia 5942  
Telephone 1300 855 585 • Fax (08) 8243 8208  
enquiries@novita.org.au • www.novitatech.org.au  
ABN 53 137 537 636

