unicare
health

TurnTable
Instruction & Safety Manual

Please read this manual before operation, for maintenance instructions and safe usage



The enabling power of applied knowledge

Introduction

The TurnTable can be used for seated transfers between wheelchair and bed, toilet or chair/armchair. For increased security, the TurnTable can be be used together with FlexiBelt/EasyBelt. TurnTable is available in three sizes with handles or without handle for tight spaces.

SystemRoMedic is a series of different models of transfer assistive devices. The choice of assistive devices depends on the transfer situation and the patient's functional capability and needs.

Functional Inspection

Visual inspection

Check the condition and function of the assistive device regularly. Check to ensure that fabric is not worn or faded. If there are signs of wear, the product must be discarded.

To transfer with TurnTable, the patient must be able to:

- support weight while standing
- stand up and maintain balance (with support)
- understand instructions

Placement and use TurnTable



Place the TurnTable on the floor beside the patient. Secure the TurnTable in place, before raising the patient, by placing the ball of your foot on the TurnTable and your heel on the floor. Steer the TurnTable by pushing with the front of the foot and braking with the heel. When the patient is about to sit down, secure the TurnTable in place as described above.



When transferring from bed to wheelchair, place the wheelchair close to the bed, so as to minimize transfer distance.



When transferring to the toilet, place the wheelchair close to the toilet, so as to minimize transfer distance. Pull the patient's pants up/down when the patient has risen slightly from the seat. Preferably, two caregivers should assist.

T 08 9351 7500 E info@unicarehealth.com.au W www.unicarehealth.com.au

Care of Product

The surface of the product can be cleaned with warm water or a soap solution. If disinfection is needed, 70% ethanol, 45% isopropanol or similar should be used.

CE

Medical Device Class 1

The product meets the requirements of the Medical Devices Directive 93/42/EEC.

Contact Unicare Health if you have any questions about the product and its use. Always make sure that you have the right version of the manual. The most recent editions of manuals are available for downloading from our website: www.unicarehealth.com.au

PO Box 82 Bentley WA 6982
T 08 9351 7500 F 08 9350 5299
info@unicarehealth.com.au
www.unicarehealth.com.au

Authorised Distributor